

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044208.D
 Acq On : 15 Jun 2021 02:24
 Operator : SY/MD
 Sample : M2687-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 17-B6(112-116)

Quant Time: Jun 15 05:00:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	174580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	319359	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	312409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	147967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	133318	48.021	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.040%	
35) Dibromofluoromethane	5.298	113	103862	46.752	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.500%	
50) Toluene-d8	7.906	98	409676	48.018	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.040%	
62) 4-Bromofluorobenzene	10.639	95	144106	42.147	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.300%	
Target Compounds						
						Qvalue
16) Acetone	2.652	43	10678	5.353	ug/l	98
20) Methylene Chloride	3.057	84	13684	4.375	ug/l	97
43) Isopropyl Acetate	5.922	43	17832	2.954	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
Data File : VU044208.D
Acq On : 15 Jun 2021 02:24
Operator : SY/MD
Sample : M2687-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
17-B6(112-116)

Quant Time: Jun 15 05:00:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

