

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044172.D
 Acq On : 14 Jun 2021 09:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 15 02:41:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.705	0.702	0.4	97	0.00
3 P	Chloromethane	0.933	0.833	10.7	95	0.00
4 C	Vinyl Chloride	0.841	0.806	4.2#	96	0.00
5 T	Bromomethane	0.347	0.453	-30.5#	112	0.00
6 T	Chloroethane	0.534	0.489	8.4	96	0.00
7 T	Trichlorofluoromethane	1.012	0.987	2.5	97	0.00
8 T	Diethyl Ether	0.414	0.399	3.6	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.610	1.1	99	0.00
10 T	Methyl Iodide	0.711	0.708	0.4	89	0.00
11 T	Tert butyl alcohol	0.214	0.194	9.3	91	-0.01
12 CM	1,1-Dichloroethene	0.615	0.591	3.9#	96	0.00
13 T	Acrolein	0.125	0.089	28.8#	71	0.00
14 T	Allyl chloride	1.147	1.138	0.8	98	0.00
15 T	Acrylonitrile	0.442	0.453	-2.5	97	0.00
16 T	Acetone	0.467	0.566	-21.2#	131	0.00
17 T	Carbon Disulfide	1.974	1.917	2.9	99	0.00
18 T	Methyl Acetate	1.023	1.066	-4.2	100	0.00
19 T	Methyl tert-butyl Ether	1.877	1.898	-1.1	95	0.00
20 T	Methylene Chloride	0.807	0.738	8.6	96	0.00
21 T	trans-1,2-Dichloroethene	0.679	0.678	0.1	98	0.00
22 T	Diisopropyl ether	1.975	2.079	-5.3	99	0.00
23 T	Vinyl Acetate	1.874	1.977	-5.5	98	0.00
24 P	1,1-Dichloroethane	1.312	1.286	2.0	97	0.00
25 T	2-Butanone	0.664	0.714	-7.5	104	0.00
26 T	2,2-Dichloropropane	1.036	1.040	-0.4	100	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.742	2.1	97	0.00
28 T	Bromochloromethane	0.626	0.650	-3.8	103	0.00
29 T	Tetrahydrofuran	0.361	0.372	-3.0	96	0.00
30 C	Chloroform	1.282	1.267	1.2#	99	0.00
31 T	Cyclohexane	1.122	1.079	3.8	96	0.00
32 T	1,1,1-Trichloroethane	1.066	1.053	1.2	98	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.810	-1.9	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.348	0.350	-0.6	97	0.00
36 T	1,1-Dichloropropene	0.556	0.558	-0.4	99	0.00
37 T	Ethyl Acetate	0.666	0.662	0.6	96	0.00
38 T	Carbon Tetrachloride	0.502	0.510	-1.6	99	0.00
39 T	Methylcyclohexane	0.593	0.580	2.2	95	0.00
40 TM	Benzene	1.648	1.641	0.4	97	0.00
41 T	Methacrylonitrile	0.341	0.362	-6.2	98	0.00
42 TM	1,2-Dichloroethane	0.582	0.574	1.4	98	0.00
43 T	Isopropyl Acetate	0.945	0.953	-0.8	96	0.00
44 TM	Trichloroethene	0.380	0.374	1.6	97	0.00
45 C	1,2-Dichloropropane	0.448	0.439	2.0#	99	0.00
46 T	Dibromomethane	0.289	0.284	1.7	98	0.00
47 T	Bromodichloromethane	0.568	0.583	-2.6	100	0.00
48 T	Methyl methacrylate	0.460	0.477	-3.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044172.D
 Acq On : 14 Jun 2021 09:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 15 02:41:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.012	0.012	0.0	90	0.00
50 S	Toluene-d8	1.336	1.335	0.1	96	0.00
51 T	4-Methyl-2-Pentanone	0.684	0.699	-2.2	95	0.00
52 CM	Toluene	0.977	1.002	-2.6#	96	0.00
53 T	t-1,3-Dichloropropene	0.621	0.642	-3.4	98	0.00
54 T	cis-1,3-Dichloropropene	0.670	0.697	-4.0	99	0.00
55 T	1,1,2-Trichloroethane	0.410	0.401	2.2	95	0.00
56 T	Ethyl methacrylate	0.627	0.647	-3.2	93	0.00
57 T	1,3-Dichloropropane	0.728	0.724	0.5	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.194	0.187	3.6	88	0.00
59 T	2-Hexanone	0.546	0.565	-3.5	96	0.00
60 T	Dibromochloromethane	0.408	0.417	-2.2	97	0.00
61 T	1,2-Dibromoethane	0.432	0.425	1.6	96	0.00
62 S	4-Bromofluorobenzene	0.535	0.523	2.2	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.344	0.343	0.3	98	0.00
65 PM	Chlorobenzene	1.069	1.029	3.7	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.368	1.3	96	0.00
67 C	Ethyl Benzene	1.920	1.961	-2.1#	96	0.00
68 T	m/p-Xylenes	0.717	0.751	-4.7	96	0.00
69 T	o-Xylene	0.680	0.694	-2.1	94	0.00
70 T	Styrene	1.185	1.235	-4.2	94	0.00
71 P	Bromoform	0.318	0.318	0.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.348	3.499	-4.5	94	0.00
74 T	N-amyl acetate	1.504	1.512	-0.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.458	1.381	5.3	93	0.00
76 T	1,2,3-Trichloropropane	1.676	1.645	1.8	94	0.00
77 T	Bromobenzene	0.842	0.805	4.4	94	0.00
78 T	n-propylbenzene	4.358	4.510	-3.5	96	0.00
79 T	2-Chlorotoluene	2.582	2.590	-0.3	94	0.00
80 T	1,3,5-Trimethylbenzene	2.979	3.175	-6.6	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.462	0.463	-0.2	94	0.00
82 T	4-Chlorotoluene	3.049	3.156	-3.5	96	0.00
83 T	tert-Butylbenzene	2.795	2.813	-0.6	92	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.217	-7.5	95	0.00
85 T	sec-Butylbenzene	3.670	3.814	-3.9	95	0.00
86 T	p-Isopropyltoluene	3.144	3.349	-6.5	94	0.00
87 T	1,3-Dichlorobenzene	1.652	1.606	2.8	94	0.00
88 T	1,4-Dichlorobenzene	1.759	1.647	6.4	94	0.00
89 T	n-Butylbenzene	3.194	3.307	-3.5	97	0.00
90 T	Hexachloroethane	0.498	0.507	-1.8	97	0.00
91 T	1,2-Dichlorobenzene	1.661	1.588	4.4	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.330	0.328	0.6	97	0.00
93 T	1,2,4-Trichlorobenzene	1.046	1.015	3.0	94	0.00
94 T	Hexachlorobutadiene	0.561	0.551	1.8	95	0.00
95 T	Naphthalene	3.306	3.254	1.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
Data File : VU044172.D
Acq On : 14 Jun 2021 09:25
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Jun 15 02:41:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	0.992	2.2	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6