

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
 Data File : VU044209.D
 Acq On : 15 Jun 2021 02:47
 Operator : SY/MD
 Sample : M2687-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 17-B6(116-120)

Quant Time: Jun 15 05:00:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.385	168	178110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.260	114	327134	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	319616	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	151047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	136016	48.021	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.040%	
35) Dibromofluoromethane	5.298	113	107301	47.152	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.300%	
50) Toluene-d8	7.906	98	422135	48.302	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.600%	
62) 4-Bromofluorobenzene	10.639	95	149321	42.634	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.260%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	3.057	84	14009	4.395	ug/l	97
43) Isopropyl Acetate	5.922	43	17234	2.787	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061421\
Data File : VU044209.D
Acq On : 15 Jun 2021 02:47
Operator : SY/MD
Sample : M2687-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
17-B6(116-120)

Quant Time: Jun 15 05:00:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.385 min Scan# 1258
Delta R.T. 0.003 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47




Instrument :
MSVOA_U
ClientSampled :
17-B6(116-120)


Tgt Ion:168 Resp: 178110
Ion Ratio Lower Upper
168 100
99 61.7 51.3 76.9



#20
Methylene Chloride
Concen: 4.395 ug/l
RT: 3.057 min Scan# 534
Delta R.T. 0.006 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Tgt Ion: 84 Resp: 14009
Ion Ratio Lower Upper
84 100
49 115.5 96.9 145.3
51 38.5 31.0 46.6
86 64.9 51.0 76.6









#33

1,2-Dichloroethane-d4
Concen: 48.021 ug/l
RT: 5.713 min Scan# 1360
Delta R.T. 0.003 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Instrument :
MSVOA_U
ClientSampleId :
17-B6(116-120)




Tgt Ion: 65 Resp: 136016
Ion Ratio Lower Upper
65 100
67 54.6 0.0 106.6




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.260 min Scan# 1530
Delta R.T. 0.003 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Tgt Ion: 114 Resp: 327134
Ion Ratio Lower Upper
114 100
63 21.4 0.0 45.8
88 16.8 0.0 34.0









#35
Dibromofluoromethane
Concen: 47.152 ug/l
RT: 5.298 min Scan# 1231
Delta R.T. 0.000 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Instrument :
MSVOA_U
ClientSampleId :
17-B6(116-120)

Tgt Ion:113 Resp: 107301




Ion	Ratio	Lower	Upper
113	100		
111	102.0	81.0	121.4
192	18.8	14.3	21.5




#43
Isopropyl Acetate
Concen: 2.787 ug/l
RT: 5.922 min Scan# 1425
Delta R.T. 0.003 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Tgt Ion: 43 Resp: 17234

Ion	Ratio	Lower	Upper
43	100		
61	22.5	17.8	26.8
87	13.3	11.1	16.7









#50
Toluene-d8
Concen: 48.302 ug/l
RT: 7.906 min Scan# 2042
Delta R.T. -0.000 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47




Instrument :
MSVOA_U
ClientSampled :
17-B6(116-120)


Tgt Ion: 98 Resp: 422135
Ion Ratio Lower Upper
98 100
100 64.6 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 42.634 ug/l
RT: 10.639 min Scan# 2892
Delta R.T. -0.000 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Tgt Ion: 95 Resp: 149321
Ion Ratio Lower Upper
95 100
174 74.7 0.0 144.4
176 70.9 0.0 138.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.423 min Scan# 2514
Delta R.T. -0.000 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Instrument :
MSVOA_U
ClientSampleId :
17-B6(116-120)

Tgt Ion:117 Resp: 319616
Ion Ratio Lower Upper
117 100
82 59.4 48.8 73.2
119 31.9 25.0 37.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.819 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VU044209.D
Acq: 15 Jun 2021 02:47

Tgt Ion:152 Resp: 151047
Ion Ratio Lower Upper
152 100
115 61.8 43.8 131.3
150 158.3 0.0 350.6

