

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061422\
 Data File : VU049141.D
 Acq On : 15 Jun 2022 02:28
 Operator : SY/MD
 Sample : N3274-07 250X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B97

Quant Time: Jun 15 07:23:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 05:38:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	135693	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	124379	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	61645	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	53515	4.513	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.200%	
7) Chloroethane-d5	1.912	69	39377	4.906	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.200%	
11) 1,1-Dichloroethene-d2	2.568	65	21615	4.532	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.600%	
20) 2-Butanone-d5	4.636	46	192968	60.954	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	121.900%	
24) Chloroform-d	5.063	84	92578	4.977	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.703	65	54118	5.303	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.000%	
32) Benzene-d6	5.729	84	190256	5.078	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.690	67	66606	5.237	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.800%	
41) Toluene-d8	7.899	98	170822	5.063	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.200%	
43) trans-1,3-Dichloroprop...	8.182	79	26772	4.739	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.800%	
46) 2-Hexanone-d5	8.636	63	123464	54.172	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.340%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	55840	5.484	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.600%	
66) 1,2-Dichlorobenzene-d4	12.195	152	61379	5.189	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.800%	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	2.138	101	930	0.070	ug/L	# 64
34) Trichloroethene	6.543	95	5715	0.604	ug/L	95
47) Tetrachloroethene	8.555	164	3209	0.492	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061422\
Data File : VU049141.D
Acq On : 15 Jun 2022 02:28
Operator : SY/MD
Sample : N3274-07 250X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B97

Quant Time: Jun 15 07:23:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 05:38:08 2022
Response via : Initial Calibration

