

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061422\
 Data File : VU049111.D
 Acq On : 14 Jun 2022 13:59
 Operator : SY/MD
 Sample : VU0614WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK085

Quant Time: Jun 15 03:16:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 03:15:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	136963	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	129648	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	63589	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	58720	4.907	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.200%
7) Chloroethane-d5	1.912	69	42089	5.195	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.800%
11) 1,1-Dichloroethene-d2	2.568	65	23189	4.817	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.400%
20) 2-Butanone-d5	4.636	46	169119	52.925	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.860%
24) Chloroform-d	5.066	84	98865	5.265	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
26) 1,2-Dichloroethane-d4	5.703	65	56959	5.530	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.600%
32) Benzene-d6	5.729	84	203372	5.207	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
36) 1,2-Dichloropropane-d6	6.693	67	69125	5.215	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.899	98	182744	5.196	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.000%
43) trans-1,3-Dichloroprop...	8.182	79	31266	5.309	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.200%
46) 2-Hexanone-d5	8.635	63	130822	55.068	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.140%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	58088	5.473	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	63235	5.182	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061422\
Data File : VU049111.D
Acq On : 14 Jun 2022 13:59
Operator : SY/MD
Sample : VU0614WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK085

Quant Time: Jun 15 03:16:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 03:15:08 2022
Response via : Initial Calibration

