

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
 Data File : VU059245.D
 Acq On : 14 Jun 2024 14:15
 Operator : MD/SY
 Sample : P2775-09ME
 Misc : 3.51g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4C87ME

Quant Time: Jun 15 02:51:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 15 02:48:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	216563	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	221669	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	100372	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	22001	16.967	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	33.940%#
7) Chloroethane-d5	1.872	69	19047	17.148	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	34.300%#
11) 1,1-Dichloroethene-d2	2.538	63	35135	13.728	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	27.460%#
21) 2-Butanone-d5	4.618	46	111205	90.989	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.990%
24) Chloroform-d	5.052	84	78305	30.336	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	60.680%#
26) 1,2-Dichloroethane-d4	5.692	65	51341	32.451	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	64.900%#
32) Benzene-d6	5.715	84	133658	26.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	52.540%#
36) 1,2-Dichloropropane-d6	6.683	67	54271	30.938	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	61.880%#
41) Toluene-d8	7.891	98	100092	23.175	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	46.360%#
43) trans-1,3-Dichloroprop...	8.178	79	19685	27.761	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.520%#
47) 2-Hexanone-d5	8.641	63	67119	82.741	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.740%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	121475	40.916	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.840%
66) 1,2-Dichlorobenzene-d4	12.190	152	54930	36.672	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.340%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
Data File : VU059245.D
Acq On : 14 Jun 2024 14:15
Operator : MD/SY
Sample : P2775-09ME
Misc : 3.51g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4C87ME

Quant Time: Jun 15 02:51:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 15 02:48:14 2024
Response via : Initial Calibration

