

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
 Data File : VU059242.D
 Acq On : 14 Jun 2024 13:10
 Operator : MD/SY
 Sample : VU0614MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK073

Quant Time: Jun 15 02:50:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 15 02:48:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	207440	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	210597	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	87613	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	66947	53.901	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	107.800%
7) Chloroethane-d5	1.907	69	61017	57.350	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.700%
11) 1,1-Dichloroethene-d2	2.560	63	89007	36.306	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.620%
21) 2-Butanone-d5	4.611	46	101709	86.879	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.880%
24) Chloroform-d	5.052	84	125913	50.924	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.840%
26) 1,2-Dichloroethane-d4	5.692	65	77723	51.287	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.580%
32) Benzene-d6	5.718	84	254958	52.748	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.500%
36) 1,2-Dichloropropane-d6	6.682	67	83742	50.248	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.500%
41) Toluene-d8	7.891	98	210211	51.231	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.460%
43) trans-1,3-Dichloroprop...	8.174	79	31093	46.155	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.320%
47) 2-Hexanone-d5	8.624	63	56857	73.775	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.780%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	120802	42.829	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.660%
66) 1,2-Dichlorobenzene-d4	12.187	152	71594	54.757	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
Data File : VU059242.D
Acq On : 14 Jun 2024 13:10
Operator : MD/SY
Sample : VU0614MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK073

Quant Time: Jun 15 02:50:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 15 02:48:14 2024
Response via : Initial Calibration

