

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
 Data File : VU059246.D
 Acq On : 14 Jun 2024 14:38
 Operator : MD/SY
 Sample : P2870-16ME
 Misc : 3.84g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB7ME

Quant Time: Jun 15 02:51:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 15 02:48:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	188774	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	195265	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	83520	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	56460	49.953	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.900%
7) Chloroethane-d5	1.907	69	52973	54.712	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.420%
11) 1,1-Dichloroethene-d2	2.560	63	73486	32.939	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.880%
21) 2-Butanone-d5	4.612	46	101377	95.158	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.160%
24) Chloroform-d	5.055	84	115496	51.330	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.660%
26) 1,2-Dichloroethane-d4	5.695	65	77275	56.033	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.060%
32) Benzene-d6	5.721	84	228835	51.061	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.120%
36) 1,2-Dichloropropane-d6	6.685	67	79595	51.509	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.020%
41) Toluene-d8	7.894	98	189536	49.819	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.640%
43) trans-1,3-Dichloroprop...	8.177	79	29530	47.277	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.560%
47) 2-Hexanone-d5	8.627	63	57911	81.043	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.040%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	118769	45.414	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.820%
66) 1,2-Dichlorobenzene-d4	12.190	152	69264	55.571	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061424\
Data File : VU059246.D
Acq On : 14 Jun 2024 14:38
Operator : MD/SY
Sample : P2870-16ME
Misc : 3.84g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB7ME

Quant Time: Jun 15 02:51:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 15 02:48:14 2024
Response via : Initial Calibration

