

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061521\
 Data File : VU044223.D
 Acq On : 15 Jun 2021 13:18
 Operator : SY/MD
 Sample : M2674-10 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TMR-DS-04-060921

Quant Time: Jun 15 13:42:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	199016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	337877	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	343550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	180568	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	140154	44.284	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.560%	
35) Dibromofluoromethane	5.298	113	109177	46.451	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.900%	
50) Toluene-d8	7.906	98	436144	48.318	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.640%	
62) 4-Bromofluorobenzene	10.642	95	176323	48.743	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.658	43	25325	13.472	ug/l	96
18) Methyl Acetate	2.967	43	6514	1.600	ug/l	100
20) Methylene Chloride	3.054	84	7474	1.456	ug/l	97
25) 2-Butanone	4.742	43	10844	4.105	ug/l	98
40) Benzene	5.780	78	11646	1.046	ug/l	99
52) Toluene	7.977	92	70555	10.686	ug/l	99
67) Ethyl Benzene	9.578	91	52802	4.001	ug/l	97
68) m/p-Xylenes	9.700	106	48486	9.838	ug/l	99
69) o-Xylene	10.108	106	31788	6.802	ug/l	98
80) 1,3,5-Trimethylbenzene	11.095	105	11944	1.110	ug/l	98
84) 1,2,4-Trimethylbenzene	11.475	105	40542	3.751	ug/l	97
95) Naphthalene	14.092	128	129790	12.482	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061521\
Data File : VU044223.D
Acq On : 15 Jun 2021 13:18
Operator : SY/MD
Sample : M2674-10 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TMR-DS-04-060921

Quant Time: Jun 15 13:42:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

