

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
 Data File : VU054515.D
 Acq On : 15 Jun 2023 10:45
 Operator : MD/SY
 Sample : 03237-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YBW67

Quant Time: Jun 16 01:30:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 14 00:05:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	295613	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	290371	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	137837	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	56901	3.025	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.400%
7) Chloroethane-d5	1.914	69	64018	3.634	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.600%
11) 1,1-Dichloroethene-d2	2.567	65	27857	3.315	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.400%
20) 2-Butanone-d5	4.628	46	248716	47.021	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.040%
24) Chloroform-d	5.059	84	160319	4.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
26) 1,2-Dichloroethane-d4	5.698	65	80664	4.208	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.200%
32) Benzene-d6	5.724	84	293251	3.633	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.600%
36) 1,2-Dichloropropane-d6	6.689	67	108636	4.030	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.600%
41) Toluene-d8	7.894	98	243303	3.396	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.000%#
43) trans-1,3-Dichloroprop...	8.177	79	34559	3.687	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.631	63	160060	40.699	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.400%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	85400	4.248	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	100888	4.339	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
Data File : VU054515.D
Acq On : 15 Jun 2023 10:45
Operator : MD/SY
Sample : 03237-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YBW67

Quant Time: Jun 16 01:30:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 14 00:05:43 2023
Response via : Initial Calibration

