

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
 Data File : VU054550.D
 Acq On : 16 Jun 2023 01:57
 Operator : MD/SY
 Sample : 03232-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONK7

Quant Time: Jun 16 02:44:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 14 00:05:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	257180	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	251395	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	121255	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	54769	3.346	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	67.000%	
7) Chloroethane-d5	1.914	69	72333	4.720	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.400%	
11) 1,1-Dichloroethene-d2	2.567	65	26730	3.657	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.200%	
20) 2-Butanone-d5	4.628	46	230622	50.116	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.240%	
24) Chloroform-d	5.058	84	161620	4.724	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
26) 1,2-Dichloroethane-d4	5.698	65	81556	4.890	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.800%	
32) Benzene-d6	5.724	84	313750	4.490	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.800%	
36) 1,2-Dichloropropane-d6	6.685	67	108189	4.636	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.800%	
41) Toluene-d8	7.894	98	249545	4.024	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.400%	
43) trans-1,3-Dichloroprop...	8.177	79	30317	3.736	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.800%	
46) 2-Hexanone-d5	8.631	63	149731	43.976	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.960%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	82274	4.727	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.600%	
66) 1,2-Dichlorobenzene-d4	12.190	152	101276	4.952	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.000%	
Target Compounds						
12) 1,1-Dichloroethene	2.579	96	26900	1.442	ug/L #	57
19) 1,1-Dichloroethane	3.865	63	103915	2.607	ug/L	98
29) 1,1,1-Trichloroethane	5.312	97	4530678	132.906	ug/L	98
34) Trichloroethene	6.537	95	116618	4.881	ug/L	93
45) 1,1,2-Trichloroethane	8.396	97	2321	0.149	ug/L #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
Data File : VU054550.D
Acq On : 16 Jun 2023 01:57
Operator : MD/SY
Sample : 03232-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONK7

Quant Time: Jun 16 02:44:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 14 00:05:43 2023
Response via : Initial Calibration

