

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
 Data File : VU054514.D
 Acq On : 15 Jun 2023 10:14
 Operator : MD/SY
 Sample : VU0615WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK002

Quant Time: Jun 16 01:30:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 14 00:05:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	269966	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	257539	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119252	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	75384	4.388	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.911	69	90930	5.653	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.000%
11) 1,1-Dichloroethene-d2	2.563	65	34648	4.515	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.400%
20) 2-Butanone-d5	4.624	46	245623	50.848	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.700%
24) Chloroform-d	5.058	84	187097	5.210	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
26) 1,2-Dichloroethane-d4	5.698	65	92899	5.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.200%
32) Benzene-d6	5.721	84	374482	5.231	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
36) 1,2-Dichloropropane-d6	6.685	67	128524	5.376	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.600%
41) Toluene-d8	7.894	98	301524	4.746	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
43) trans-1,3-Dichloroprop...	8.177	79	36779	4.424	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.400%
46) 2-Hexanone-d5	8.630	63	163024	46.738	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.480%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	90797	5.093	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	107527	5.346	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
Data File : VU054514.D
Acq On : 15 Jun 2023 10:14
Operator : MD/SY
Sample : VU0615WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK002

Quant Time: Jun 16 01:30:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 14 00:05:43 2023
Response via : Initial Calibration

