

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
 Data File : VU054554.D
 Acq On : 16 Jun 2023 03:34
 Operator : MD/SY
 Sample : 03232-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONQ9

Quant Time: Jun 16 04:25:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 14 00:05:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	256995	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	252176	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	121580	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	54288	3.319	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.400%
7) Chloroethane-d5	1.914	69	72688	4.747	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.567	65	26235	3.591	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.800%
20) 2-Butanone-d5	4.631	46	229064	49.813	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.620%
24) Chloroform-d	5.058	84	165963	4.855	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.698	65	81840	4.911	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.724	84	311669	4.446	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.685	67	108341	4.628	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.894	98	253627	4.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.600%
43) trans-1,3-Dichloroprop...	8.177	79	30629	3.762	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.631	63	149159	43.672	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.340%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	83369	4.776	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	99864	4.870	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%
Target Compounds						
22) cis-1,2-Dichloroethene	4.663	96	3707	0.169	ug/L	88
34) Trichloroethene	6.541	95	13885	0.579	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061523\
Data File : VU054554.D
Acq On : 16 Jun 2023 03:34
Operator : MD/SY
Sample : 03232-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONQ9

Quant Time: Jun 16 04:25:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 14 00:05:43 2023
Response via : Initial Calibration









#22
 cis-1,2-Dichloroethene
 Concen: 0.169 ug/L
 RT: 4.663 min Scan# 1049
 Delta R.T. -0.000 min
 Lab File: VU054554.D
 Acq: 16 Jun 2023 03:34

Instrument :
 MSVOA_U
 ClientSampleId :
 CONQ9

Tgt Ion: 96 Resp: 3707
 Ion Ratio Lower Upper
 96 100
 61 121.1 96.5 179.1
 98 57.1 44.9 83.3



#34
 Trichloroethene
 Concen: 0.579 ug/L
 RT: 6.541 min Scan# 1633
 Delta R.T. -0.000 min
 Lab File: VU054554.D
 Acq: 16 Jun 2023 03:34

Tgt Ion: 95 Resp: 13885
 Ion Ratio Lower Upper
 95 100
 97 64.2 46.8 87.0
 132 89.8 68.4 127.0
 130 98.7 68.7 127.5

