

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061620\
 Data File : VU038923.D
 Acq On : 16 Jun 2020 15:08
 Operator : SY/MD
 Sample : VIBLK21
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK21

Quant Time: Jun 17 02:47:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061520WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 17 02:00:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	286558	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	289160	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	135221	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	87210	6.11	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.20%
7) Chloroethane-d5	1.93	69	71649	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.59	63	138399	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	4.68	46	207351	40.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.24%
24) Chloroform-d	5.10	84	172498	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.74	65	103337	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.00%
32) Benzene-d6	5.76	84	343167	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
36) 1,2-Dichloropropane-d6	6.72	67	105035	5.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.60%
41) Toluene-d8	7.92	98	293558	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	8.20	79	43388	4.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.80%
46) 2-Hexanone-d5	8.66	63	244309	55.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.76%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	99784	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	119254	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061620\
Data File : VU038923.D
Acq On : 16 Jun 2020 15:08
Operator : SY/MD
Sample : VIBLK21
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK21

Quant Time: Jun 17 02:47:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061520WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 17 02:00:04 2020
Response via : Initial Calibration

