

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
 Data File : VU044236.D
 Acq On : 16 Jun 2021 13:25
 Operator : SY/MD
 Sample : M2714-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 33242

Quant Time: Jun 17 04:56:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	225655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	376717	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	367396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	196877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	150506	41.941	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.880%	
35) Dibromofluoromethane	5.298	113	121185	46.244	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.480%	
50) Toluene-d8	7.906	98	480146	47.709	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.420%	
62) 4-Bromofluorobenzene	10.639	95	187249	46.427	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.860%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	3.057	84	8411	1.436	ug/l	94
25) 2-Butanone	4.732	43	23459	7.832	ug/l	96
95) Naphthalene	14.092	128	10364	3.542	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
Data File : VU044236.D
Acq On : 16 Jun 2021 13:25
Operator : SY/MD
Sample : M2714-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
33242

Quant Time: Jun 17 04:56:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

