

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
 Data File : VU044239.D
 Acq On : 16 Jun 2021 14:34
 Operator : SY/MD
 Sample : M2714-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 33242

Quant Time: Jun 17 04:57:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	234924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	393594	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	377344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	204255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	154822	41.442	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.880%	
35) Dibromofluoromethane	5.298	113	126125	46.065	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.140%	
50) Toluene-d8	7.906	98	493439	46.927	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.860%	
62) 4-Bromofluorobenzene	10.639	95	193214	45.852	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.700%	
Target Compounds						
20) Methylene Chloride	3.057	84	8805	1.451	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
Data File : VU044239.D
Acq On : 16 Jun 2021 14:34
Operator : SY/MD
Sample : M2714-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
33242

Quant Time: Jun 17 04:57:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

