

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
 Data File : VU044240.D
 Acq On : 16 Jun 2021 14:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 04:57:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	137	0.00
2 T	Dichlorodifluoromethane	0.705	0.545	22.7#	110	0.00
3 P	Chloromethane	0.933	0.607	34.9#	102	0.00
4 C	Vinyl Chloride	0.841	0.643	23.5#	112	0.00
5 T	Bromomethane	0.347	0.319	8.1	115	0.03
6 T	Chloroethane	0.534	0.316	40.8#	91	0.01
7 T	Trichlorofluoromethane	1.012	0.806	20.4#	116	0.00
8 T	Diethyl Ether	0.414	0.345	16.7	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.522	15.4	125	0.00
10 T	Methyl Iodide	0.711	0.688	3.2	127	0.00
11 T	Tert butyl alcohol	0.214	0.156	27.1#	108	0.05
12 CM	1,1-Dichloroethene	0.615	0.530	13.8#	126	0.00
13 T	Acrolein	0.125	0.056	55.2#	66	0.00
14 T	Allyl chloride	1.147	0.921	19.7	116	0.00
15 T	Acrylonitrile	0.442	0.379	14.3	119	0.00
16 T	Acetone	0.467	0.339	27.4#	115	0.01
17 T	Carbon Disulfide	1.974	1.561	20.9#	118	0.00
18 T	Methyl Acetate	1.023	0.894	12.6	122	0.00
19 T	Methyl tert-butyl Ether	1.877	1.713	8.7	126	0.00
20 T	Methylene Chloride	0.807	0.648	19.7	123	0.00
21 T	trans-1,2-Dichloroethene	0.679	0.589	13.3	124	0.00
22 T	Diisopropyl ether	1.975	1.630	17.5	114	0.00
23 T	Vinyl Acetate	1.874	1.566	16.4	113	0.00
24 P	1,1-Dichloroethane	1.312	1.088	17.1	121	0.00
25 T	2-Butanone	0.664	0.544	18.1	116	0.00
26 T	2,2-Dichloropropane	1.036	0.878	15.3	124	0.00
27 T	cis-1,2-Dichloroethene	0.758	0.681	10.2	130	0.00
28 T	Bromochloromethane	0.626	0.499	20.3#	115	0.00
29 T	Tetrahydrofuran	0.361	0.312	13.6	117	0.00
30 C	Chloroform	1.282	1.077	16.0#	122	0.00
31 T	Cyclohexane	1.122	0.896	20.1#	117	0.00
32 T	1,1,1-Trichloroethane	1.066	0.910	14.6	124	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.634	20.3#	113	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	134	0.00
35 S	Dibromofluoromethane	0.348	0.318	8.6	124	0.00
36 T	1,1-Dichloropropene	0.556	0.491	11.7	122	0.00
37 T	Ethyl Acetate	0.666	0.585	12.2	120	0.00
38 T	Carbon Tetrachloride	0.502	0.454	9.6	124	0.00
39 T	Methylcyclohexane	0.593	0.549	7.4	127	0.00
40 TM	Benzene	1.648	1.462	11.3	122	0.00
41 T	Methacrylonitrile	0.341	0.304	10.9	116	0.00
42 TM	1,2-Dichloroethane	0.582	0.480	17.5	116	0.00
43 T	Isopropyl Acetate	0.945	0.856	9.4	121	0.00
44 TM	Trichloroethene	0.380	0.371	2.4	135	0.00
45 C	1,2-Dichloropropane	0.448	0.380	15.2#	120	0.00
46 T	Dibromomethane	0.289	0.255	11.8	124	0.00
47 T	Bromodichloromethane	0.568	0.498	12.3	121	0.00
48 T	Methyl methacrylate	0.460	0.409	11.1	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
 Data File : VU044240.D
 Acq On : 16 Jun 2021 14:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 04:57:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 05:11:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.012	0.010	16.7	107	0.04
50 S	Toluene-d8	1.336	1.190	10.9	121	0.00
51 T	4-Methyl-2-Pentanone	0.684	0.595	13.0	114	0.00
52 CM	Toluene	0.977	0.914	6.4#	124	0.00
53 T	t-1,3-Dichloropropene	0.621	0.581	6.4	125	0.00
54 T	cis-1,3-Dichloropropene	0.670	0.624	6.9	124	0.00
55 T	1,1,2-Trichloroethane	0.410	0.367	10.5	123	0.00
56 T	Ethyl methacrylate	0.627	0.613	2.2	124	0.00
57 T	1,3-Dichloropropane	0.728	0.645	11.4	122	0.00
58 T	2-Chloroethyl Vinyl ether	0.194	0.205	-5.7	135	0.00
59 T	2-Hexanone	0.546	0.463	15.2	111	0.00
60 T	Dibromochloromethane	0.408	0.386	5.4	127	0.00
61 T	1,2-Dibromoethane	0.432	0.399	7.6	127	0.00
62 S	4-Bromofluorobenzene	0.535	0.487	9.0	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	131	0.00
64 T	Tetrachloroethene	0.344	0.345	-0.3	135	0.00
65 PM	Chlorobenzene	1.069	1.010	5.5	128	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.353	5.4	126	0.00
67 C	Ethyl Benzene	1.920	1.871	2.6#	125	0.00
68 T	m/p-Xylenes	0.717	0.713	0.6	125	0.00
69 T	o-Xylene	0.680	0.688	-1.2	127	0.00
70 T	Styrene	1.185	1.182	0.3	124	0.00
71 P	Bromoform	0.318	0.311	2.2	128	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
73 T	Isopropylbenzene	3.348	3.393	-1.3	126	0.00
74 T	N-amyl acetate	1.504	1.426	5.2	120	0.00
75 P	1,1,2,2-Tetrachloroethane	1.458	1.255	13.9	116	0.00
76 T	1,2,3-Trichloropropane	1.676	1.496	10.7	117	0.00
77 T	Bromobenzene	0.842	0.805	4.4	129	0.00
78 T	n-propylbenzene	4.358	4.165	4.4	122	0.00
79 T	2-Chlorotoluene	2.582	2.395	7.2	120	0.00
80 T	1,3,5-Trimethylbenzene	2.979	2.962	0.6	122	0.00
81 T	trans-1,4-Dichloro-2-butene	0.462	0.440	4.8	122	0.00
82 T	4-Chlorotoluene	3.049	2.858	6.3	120	0.00
83 T	tert-Butylbenzene	2.795	2.807	-0.4	125	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.049	-1.9	123	0.00
85 T	sec-Butylbenzene	3.670	3.646	0.7	124	0.00
86 T	p-Isopropyltoluene	3.144	3.241	-3.1	125	0.00
87 T	1,3-Dichlorobenzene	1.652	1.574	4.7	127	0.00
88 T	1,4-Dichlorobenzene	1.759	1.584	9.9	125	0.00
89 T	n-Butylbenzene	3.194	3.108	2.7	125	0.00
90 T	Hexachloroethane	0.498	0.460	7.6	121	0.00
91 T	1,2-Dichlorobenzene	1.661	1.538	7.4	125	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.330	0.305	7.6	123	0.00
93 T	1,2,4-Trichlorobenzene	1.046	1.116	-6.7	142	0.00
94 T	Hexachlorobutadiene	0.561	0.569	-1.4	135	0.00
95 T	Naphthalene	3.306	3.653	-10.5	143	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061621\
Data File : VU044240.D
Acq On : 16 Jun 2021 14:59
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Jun 17 04:57:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060921W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 05:11:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	1.065	-5.0	140	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6