

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061625\
 Data File : VU063405.D
 Acq On : 16 Jun 2025 15:16
 Operator : MD/SY
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5112

Quant Time: Jun 16 16:26:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 13:24:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	13984	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	19649	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	5883	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	6300	0.510	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	5.709	65	3435	0.505	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	6.687	67	4011	0.522	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.887	98	8247	0.501	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3285	0.543	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	108.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	11712	0.513	ug/L	99
6) Trichloroethene	6.544	95	6375	0.491	ug/L	95
9) 1,2-Dibromoethane	8.924	107	4460	0.538	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	4199	0.535	ug/L	94
13) 1,2-Dibromo-3-chloropr...	12.990	75	713	0.514	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061625\
Data File : VU063405.D
Acq On : 16 Jun 2025 15:16
Operator : MD/SY
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5112

Quant Time: Jun 16 16:26:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 13:24:10 2025
Response via : Initial Calibration

