

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061625\
 Data File : VU063402.D
 Acq On : 16 Jun 2025 12:54
 Operator : MD/SY
 Sample : VSTDICV0.5
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV061

Quant Time: Jun 16 13:32:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 13:24:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	14010	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	19668	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	5817	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	6437	0.520	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	3545	0.520	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	4145	0.539	ug/L	0.00
8) Toluene-d8	7.887	98	8605	0.522	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3475	0.574	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	11687	0.511	ug/L	99
6) Trichloroethene	6.544	95	6471	0.498	ug/L	96
9) 1,2-Dibromoethane	8.924	107	4505	0.543	ug/L	98
12) 1,2,3-Trichloropropane	10.814	75	4372	0.564	ug/L	94
13) 1,2-Dibromo-3-chloropr...	12.990	75	847	0.617	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061625\
Data File : VU063402.D
Acq On : 16 Jun 2025 12:54
Operator : MD/SY
Sample : VSTDICV0.5
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV061

Quant Time: Jun 16 13:32:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 13:24:10 2025
Response via : Initial Calibration

