

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061719\
 Data File : VU032787.D
 Acq On : 17 Jun 2019 17:04
 Operator : JC/SP
 Sample : K3363-06ME
 Misc : 3.25g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0JW9ME

Quant Time: Jun 18 04:01:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 18 02:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.87	114	235379	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	227049	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	124658	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	47605	29.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	59.18%#
7) Chloroethane-d5	1.64	69	28224	22.00	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	44.00%#
11) 1,1-Dichloroethene-d2	2.22	63	79046	24.39	ug/L	-0.05
Spiked Amount	50.000	Range	60 - 125	Recovery	=	48.78%#
21) 2-Butanone-d5	4.19	46	101291	75.65	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.65%
24) Chloroform-d	4.64	84	98637	31.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	63.08%#
26) 1,2-Dichloroethane-d4	5.30	65	69828	33.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.04%#
32) Benzene-d6	5.32	84	200123	33.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.62%#
36) 1,2-Dichloropropane-d6	6.32	67	65160	34.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	69.12%#
41) Toluene-d8	7.56	98	183660	33.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	66.28%#
43) trans-1,3-Dichloropropene-	7.85	79	31905	32.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.06%
47) 2-Hexanone-d5	8.32	63	69421	78.85	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.85%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	101372	34.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	69.56%
64) 1,2-Dichlorobenzene-d4	11.86	152	81414	33.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.50%#

Target Compounds

					Ovalue
13) Acetone	2.35	43	2506	1.355 ug/L	82
15) Methyl Acetate	2.63	43	14381	6.860 ug/L	98
20) cis-1,2-Dichloroethene	4.21	96	4076	2.328 ug/L	100
34) Trichloroethene	6.17	95	197758	113.161 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061719\
Data File : VU032787.D
Acq On : 17 Jun 2019 17:04
Operator : JC/SP
Sample : K3363-06ME
Misc : 3.25µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JW9ME

Quant Time: Jun 18 04:01:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 18 02:39:23 2019
Response via : Initial Calibration

