

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061719\
 Data File : VU032779.D
 Acq On : 17 Jun 2019 14:01
 Operator : JC/SP
 Sample : VU0617WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK32

Quant Time: Jun 18 03:01:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 18 02:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	227408	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	216524	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	109332	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	69214	44.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.08%
7) Chloroethane-d5	1.68	69	57911	46.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.44%
11) 1,1-Dichloroethene-d2	2.27	63	106396	33.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.96%
21) 2-Butanone-d5	4.17	46	125725	97.19	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.19%
24) Chloroform-d	4.64	84	138493	45.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.68%
26) 1,2-Dichloroethane-d4	5.30	65	95323	47.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.74%
32) Benzene-d6	5.33	84	272994	48.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.74%
36) 1,2-Dichloropropane-d6	6.32	67	87923	48.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.80%
41) Toluene-d8	7.56	98	253301	47.92	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.84%
43) trans-1,3-Dichloropropene-	7.84	79	43715	46.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.48%
47) 2-Hexanone-d5	8.31	63	82095	97.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.78%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	126467	45.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	102763	48.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061719\
Data File : VU032779.D
Acq On : 17 Jun 2019 14:01
Operator : JC/SP
Sample : VU0617WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK32

Quant Time: Jun 18 03:01:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 18 02:39:23 2019
Response via : Initial Calibration

