

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061719\
 Data File : VU032786.D
 Acq On : 17 Jun 2019 16:41
 Operator : JC/SP
 Sample : K3383-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: Jun 18 03:59:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 18 02:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	219702	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	209954	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	105795	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	64835	43.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.36%
7) Chloroethane-d5	1.68	69	54069	45.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.30%
11) 1,1-Dichloroethene-d2	2.27	63	101262	33.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.96%
21) 2-Butanone-d5	4.17	46	116188	92.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.97%
24) Chloroform-d	4.64	84	131664	45.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.22%
26) 1,2-Dichloroethane-d4	5.30	65	89876	46.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.46%
32) Benzene-d6	5.33	84	257190	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
36) 1,2-Dichloropropane-d6	6.32	67	83845	48.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.18%
41) Toluene-d8	7.56	98	235537	45.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.90%
43) trans-1,3-Dichloropropene-	7.84	79	39888	43.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.96%
47) 2-Hexanone-d5	8.31	63	77000	94.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.58%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	121924	45.24	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.48%
64) 1,2-Dichlorobenzene-d4	11.85	152	98045	47.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.78%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061719\
Data File : VU032786.D
Acq On : 17 Jun 2019 16:41
Operator : JC/SP
Sample : K3383-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: Jun 18 03:59:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 18 02:39:23 2019
Response via : Initial Calibration

