

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061720\
 Data File : VU038944.D
 Acq On : 17 Jun 2020 15:21
 Operator : SY/MD
 Sample : L3040-16
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9L69

Quant Time: Jun 18 04:30:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 18 04:07:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	344958	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	342295	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	164832	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	83428	3.48	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.60%
7) Chloroethane-d5	1.93	69	91740	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.59	63	133006	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	4.67	46	292801	38.21	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.42%
24) Chloroform-d	5.10	84	184168	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
26) 1,2-Dichloroethane-d4	5.74	65	110394	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
32) Benzene-d6	5.76	84	355939	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.60%
36) 1,2-Dichloropropane-d6	6.72	67	111730	4.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.80%
41) Toluene-d8	7.92	98	300820	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.00%
43) trans-1,3-Dichloropropene-	8.20	79	48844	3.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.00%
46) 2-Hexanone-d5	8.66	63	298766	44.74	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.48%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	101902	3.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	127160	4.13	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061720\
Data File : VU038944.D
Acq On : 17 Jun 2020 15:21
Operator : SY/MD
Sample : L3040-16
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9L69

Quant Time: Jun 18 04:30:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 18 04:07:55 2020
Response via : Initial Calibration

