

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061722\
 Data File : VU049244.D
 Acq On : 17 Jun 2022 14:18
 Operator : SY/MD
 Sample : N3277-16DL 50000X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BB5DL

Quant Time: Jun 18 04:13:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	139172	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	128386	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	62256	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	54193	4.456	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.200%
7) Chloroethane-d5	1.912	69	38191	4.639	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	2.568	65	21232	4.341	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	4.636	46	162703	50.109	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.220%
24) Chloroform-d	5.063	84	87694	4.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.703	65	49698	4.748	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	5.729	84	178372	4.612	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
36) 1,2-Dichloropropane-d6	6.693	67	60876	4.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.899	98	157900	4.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
43) trans-1,3-Dichloroprop...	8.179	79	23182	3.975	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.636	63	114176	48.534	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.060%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	48643	4.628	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	57528	4.816	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.400%
Target Compounds						
9) Trichlorofluoromethane	2.134	101	1825	0.134	ug/L	94
34) Trichloroethene	6.542	95	138869	14.221	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061722\
Data File : VU049244.D
Acq On : 17 Jun 2022 14:18
Operator : SY/MD
Sample : N3277-16DL 50000X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BB5DL

Quant Time: Jun 18 04:13:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 04:11:00 2022
Response via : Initial Calibration

