

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061722\
 Data File : VU049256.D
 Acq On : 17 Jun 2022 20:07
 Operator : SY/MD
 Sample : N3320-20DL 20X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGJ95DL

Quant Time: Jun 18 05:49:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 05:20:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	128768	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	119647	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	58066	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	49932	4.438	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.800%	
7) Chloroethane-d5	1.912	69	34708	4.556	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.200%	
11) 1,1-Dichloroethene-d2	2.565	65	19027	4.204	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.000%	
20) 2-Butanone-d5	4.633	46	159431	53.069	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.140%	
24) Chloroform-d	5.063	84	82603	4.679	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.600%	
26) 1,2-Dichloroethane-d4	5.703	65	47049	4.859	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.729	84	164114	4.553	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.000%	
36) 1,2-Dichloropropane-d6	6.690	67	56869	4.649	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.000%	
41) Toluene-d8	7.899	98	144213	4.443	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.800%	
43) trans-1,3-Dichloroprop...	8.179	79	20121	3.702	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.000%	
46) 2-Hexanone-d5	8.636	63	117528	53.607	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	107.220%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	47793	4.879	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.600%	
66) 1,2-Dichlorobenzene-d4	12.195	152	54171	4.862	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.200%	
Target Compounds						
12) 1,1-Dichloroethene	2.581	96	8166	1.128	ug/L #	5
19) 1,1-Dichloroethane	3.871	63	3855	0.244	ug/L #	90
22) cis-1,2-Dichloroethene	4.671	96	1722	0.192	ug/L	96
34) Trichloroethene	6.542	95	86992	9.559	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061722\
Data File : VU049256.D
Acq On : 17 Jun 2022 20:07
Operator : SY/MD
Sample : N3320-20DL 20X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGJ95DL

Quant Time: Jun 18 05:49:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 05:20:12 2022
Response via : Initial Calibration

