

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061725\
 Data File : VU063408.D
 Acq On : 17 Jun 2025 10:11
 Operator : MD/SY
 Sample : VU0617WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK024

Quant Time: Jun 17 10:43:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 13:24:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	13629	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	19272	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	5530	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	6475	0.538	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.000%
4) 1,2-Dichloroethane-d4	5.709	65	3489	0.526	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	6.687	67	4152	0.551	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.000%
8) Toluene-d8	7.887	98	8536	0.529	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3153	0.532	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	106.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061725\
Data File : VU063408.D
Acq On : 17 Jun 2025 10:11
Operator : MD/SY
Sample : VU0617WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK024

Quant Time: Jun 17 10:43:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 13:24:10 2025
Response via : Initial Calibration

