

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024689.D
 Acq On : 18 Jun 2018 14:11
 Operator : MD/SY
 Sample : J3463-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T2-MW-1-9.0-061218

Quant Time: Jun 18 18:25:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	322454	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	300936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	151099	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	142171	42.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.94%	
35) Dibromofluoromethane	4.89	113	128676	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	7.57	98	465572	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
62) 4-Bromofluorobenzene	10.31	95	164557	42.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	186128	104.69	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024689.D
Acq On : 18 Jun 2018 14:11
Operator : MD/SY
Sample : J3463-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T2-MW-1-9.0-061218

Quant Time: Jun 18 18:25:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

