

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024691.D
 Acq On : 18 Jun 2018 15:00
 Operator : MD/SY
 Sample : J3463-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-MW-13-8.0-061218

Quant Time: Jun 18 19:00:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	318432	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	294429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	154721	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	140949	43.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.14%	
35) Dibromofluoromethane	4.89	113	126852	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	7.57	98	461599	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
62) 4-Bromofluorobenzene	10.31	95	164489	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
Target Compounds						
16) Acetone	2.32	43	84435	48.02	ug/l	96
19) Methyl tert-butyl Ether	3.00	73	147591	17.20	ug/l	99
40) Benzene	5.40	78	16326	1.53	ug/l	100
73) Isopropylbenzene	10.17	105	61548	5.18	ug/l	99
78) n-propylbenzene	10.59	91	79007	5.65	ug/l	98
85) sec-Butylbenzene	11.33	105	33919	2.74	ug/l	97
95) Naphthalene	13.75	128	158404	14.41	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024691.D
Acq On : 18 Jun 2018 15:00
Operator : MD/SY
Sample : J3463-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-MW-13-8.0-061218

Quant Time: Jun 18 19:00:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

