

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024696.D
 Acq On : 18 Jun 2018 17:01
 Operator : MD/SY
 Sample : J3463-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-MW-12-9.0-061218

Quant Time: Jun 19 06:49:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	194102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	312812	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	287612	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	150341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	140904	44.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.94%	
35) Dibromofluoromethane	4.89	113	122699	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
50) Toluene-d8	7.57	98	449398	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
62) 4-Bromofluorobenzene	10.31	95	159949	42.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.08%	
Target Compounds						
11) Tert butyl alcohol	2.83	59	16478	21.38	ug/l	94
16) Acetone	2.32	43	24801	14.57	ug/l	95
19) Methyl tert-butyl Ether	3.00	73	297586	35.82	ug/l	96
73) Isopropylbenzene	10.17	105	7402	0.64	ug/l	94
78) n-propylbenzene	10.59	91	12396	0.91	ug/l	97
85) sec-Butylbenzene	11.33	105	3979	0.33	ug/l	99
95) Naphthalene	13.75	128	24457	2.86	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024696.D
Acq On : 18 Jun 2018 17:01
Operator : MD/SY
Sample : J3463-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-MW-12-9.0-061218

Quant Time: Jun 19 06:49:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

