

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024701.D
 Acq On : 18 Jun 2018 19:02
 Operator : MD/SY
 Sample : J3508-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KGS-SIDE-1

Quant Time: Jun 19 07:13:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	193906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	306719	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	286924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	144539	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	134983	42.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.28%	
35) Dibromofluoromethane	4.89	113	120696	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
50) Toluene-d8	7.57	98	446387	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	10.31	95	156765	42.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.04%	
Target Compounds						
16) Acetone	2.32	43	12201	7.17	ug/l	95
39) Methylcyclohexane	6.42	83	1978	0.44	ug/l	97
52) Toluene	7.64	92	41507	6.28	ug/l	96
67) Ethyl Benzene	9.26	91	30453	2.37	ug/l	99
68) m/p-Xylenes	9.38	106	30569	6.20	ug/l	89
69) o-Xylene	9.78	106	12911	2.64	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024701.D
Acq On : 18 Jun 2018 19:02
Operator : MD/SY
Sample : J3508-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KGS-SIDE-1

Quant Time: Jun 19 07:13:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

