

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024714.D
 Acq On : 19 Jun 2018 01:30
 Operator : MD/SY
 Sample : J3535-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 Z2TB02-180613

Quant Time: Jun 19 09:48:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	188282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	299524	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	274377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147469	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	132045	42.96	ug/l	0.00
Spiked Amount			Recovery	=	85.92%	
35) Dibromofluoromethane	4.89	113	116096	46.70	ug/l	0.00
Spiked Amount			Recovery	=	93.40%	
50) Toluene-d8	7.57	98	432976	47.75	ug/l	0.00
Spiked Amount			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	10.31	95	153524	42.64	ug/l	0.00
Spiked Amount			Recovery	=	85.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024714.D
Acq On : 19 Jun 2018 01:30
Operator : MD/SY
Sample : J3535-20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
Z2TB02-180613

Quant Time: Jun 19 09:48:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

