

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024718.D
 Acq On : 19 Jun 2018 03:07
 Operator : MD/SY
 Sample : J3535-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KY070MW236-180613

Quant Time: Jun 19 08:03:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	300032	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	150566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	133282	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
35) Dibromofluoromethane	4.89	113	117426	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
50) Toluene-d8	7.57	98	436581	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
62) 4-Bromofluorobenzene	10.31	95	155638	43.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.32%	
Target Compounds						
4) Vinyl Chloride	1.40	62	5646	2.25	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	3330	1.21	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024718.D
Acq On : 19 Jun 2018 03:07
Operator : MD/SY
Sample : J3535-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY070MW236-180613

Quant Time: Jun 19 08:03:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

