

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024721.D
 Acq On : 19 Jun 2018 04:20
 Operator : MD/SY
 Sample : J3535-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0448-180613

Quant Time: Jun 19 08:18:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	299636	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	278624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	149966	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	134124	43.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.30%	
35) Dibromofluoromethane	4.89	113	118251	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	7.57	98	437906	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	10.31	95	156175	43.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	5282	1.92	ug/l	80
64) Tetrachloroethene	8.23	164	3389	1.28	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024721.D
Acq On : 19 Jun 2018 04:20
Operator : MD/SY
Sample : J3535-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0448-180613

Quant Time: Jun 19 08:18:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

