

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061818\
 Data File : VU024727.D
 Acq On : 19 Jun 2018 06:45
 Operator : MD/SY
 Sample : J3535-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS042RW181-180614

Quant Time: Jun 19 08:43:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190887	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	301519	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	149407	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	134882	43.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.56%	
35) Dibromofluoromethane	4.89	113	116900	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
50) Toluene-d8	7.57	98	437498	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
62) 4-Bromofluorobenzene	10.31	95	155580	42.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.86%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061818\
Data File : VU024727.D
Acq On : 19 Jun 2018 06:45
Operator : MD/SY
Sample : J3535-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS042RW181-180614

Quant Time: Jun 19 08:43:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 4.99 min Scan# 1368

Delta R.T. -0.00 min

Lab File: VU024727.D

Acq: 19 Jun 2018 06:45

Instrument :

MSVOA_U

ClientSampleId :


SS042RW181-180614

Tot Ion:168 Resp: 190887

Ion Ratio Lower Upper




168 100

99 52.6 43.2 64.8



Abundance Ion 167.90 (167.60 to 168.60): VU024727.D (-1351) (-)

Ion 98.90 (98.60 to 99.60): VU024727.D (-1351) (-)



#33

1,2-Dichloroethane-d4

Concen: 43.28 ug/l

RT: 5.31 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VU024727.D


Acq: 19 Jun 2018 06:45

Tgt Ion: 65 Resp: 134882

Ion Ratio Lower Upper



65 100


67 53.0 0.0 107.0



Abundance Ion 64.90 (64.60 to 65.60): VU024727.D (-1451) (-)

Ion 66.90 (66.60 to 67.60): VU024727.D (-1451) (-)





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 5.89 min Scan# 1649

Delta R.T. -0.00 min

Lab File: VU024727.D

Acq: 19 Jun 2018 06:45

Instrument :

MSVOA_U

ClientSampleId :

SS042RW181-180614


Tot Ion:114 Resp: 301519

Ion Ratio Lower Upper

114 100

63 19.8 0.0 45.4




88 15.5 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VU0

Ion 87.90 (87.60 to 88.60): VU0



#35

Dibromofluoromethane

Concen: 46.72 ug/l

RT: 4.89 min Scan# 1337

Delta R.T. 0.00 min

Lab File: VU024727.D

Acq: 19 Jun 2018 06:45


Tgt Ion:113 Resp: 116900

Ion Ratio Lower Upper

113 100

111 102.2 82.2 123.4



192 21.0 16.2 24.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.93 ug/l

RT: 7.57 min Scan# 2171

Delta R.T. -0.00 min

Lab File: VU024727.D

Acq: 19 Jun 2018 06:45

Instrument :

MSVOA_U

ClientSampleId :


SS042RW181-180614

Tot Ion: 98 Resp: 437498

Ion Ratio Lower Upper




98 100

100 63.9 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VU0

Ion 100.00 (99.70 to 100.70): VU0



#62

4-Bromofluorobenzene

Concen: 42.93 ug/l

RT: 10.31 min Scan# 3023

Delta R.T. -0.00 min

Lab File: VU024727.D

Acq: 19 Jun 2018 06:45


Tgt Ion: 95 Resp: 155580

Ion Ratio Lower Upper

95 100

174 87.2 0.0 165.8



176 85.1 0.0 159.4




Abundance Ion 94.90 (94.60 to 95.60): VU0

Ion 173.80 (173.50 to 174.50): VU0

Ion 175.80 (175.50 to 176.50): VU0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 9.09 min Scan# 2645
Delta R.T. -0.00 min
Lab File: VU024727.D
Acq: 19 Jun 2018 06:45

Instrument :
MSVOA_U
ClientSampleId :
SS042RW181-180614

Tot Ion	Ratio	Lower	Upper
117	100		
82	54.0	44.3	66.5
119	32.9	25.4	38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 11.49 min Scan# 3389
Delta R.T. -0.00 min
Lab File: VU024727.D
Acq: 19 Jun 2018 06:45

Tgt Ion	Ratio	Lower	Upper
152	100		
115	56.0	43.0	129.0
150	155.9	0.0	354.0

