

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061819\
 Data File : VU032810.D
 Acq On : 18 Jun 2019 09:40
 Operator : JC/SP
 Sample : VU0618WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK34

Quant Time: Jun 19 01:41:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 01:26:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	222384	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	217623	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	110803	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	60134	39.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.14%
7) Chloroethane-d5	1.68	69	52031	42.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.86%
11) 1,1-Dichloroethene-d2	2.27	63	95626	31.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.46%
21) 2-Butanone-d5	4.17	46	121064	95.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.70%
24) Chloroform-d	4.64	84	129757	43.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.84%
26) 1,2-Dichloroethane-d4	5.30	65	91381	46.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.86%
32) Benzene-d6	5.33	84	253296	44.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.30%
36) 1,2-Dichloropropane-d6	6.32	67	83002	45.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.86%
41) Toluene-d8	7.56	98	233593	43.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.94%
43) trans-1,3-Dichloropropene-	7.85	79	41881	44.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.10%
47) 2-Hexanone-d5	8.31	63	78613	93.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	124217	44.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.92%
64) 1,2-Dichlorobenzene-d4	11.86	152	99886	46.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061819\
Data File : VU032810.D
Acq On : 18 Jun 2019 09:40
Operator : JC/SP
Sample : VU0618WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK34

Quant Time: Jun 19 01:41:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 01:26:44 2019
Response via : Initial Calibration

