

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061819\
 Data File : VU032838.D
 Acq On : 18 Jun 2019 21:19
 Operator : JC/SP
 Sample : K3324-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: Jun 19 06:41:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 04:35:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	219502	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	212135	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	107842	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	57667	38.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.88%
7) Chloroethane-d5	1.68	69	51860	43.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.70%
11) 1,1-Dichloroethene-d2	2.27	63	90977	30.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.20%
21) 2-Butanone-d5	4.17	46	127319	101.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.97%
24) Chloroform-d	4.64	84	129751	44.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.98%
26) 1,2-Dichloroethane-d4	5.30	65	91974	47.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.70%
32) Benzene-d6	5.33	84	250272	45.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.52%
36) 1,2-Dichloropropane-d6	6.32	67	82487	46.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.64%
41) Toluene-d8	7.56	98	225977	43.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.28%
43) trans-1,3-Dichloropropene-	7.84	79	40149	43.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.62%
47) 2-Hexanone-d5	8.31	63	82829	100.70	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.70%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	126364	46.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	97319	46.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061819\
Data File : VU032838.D
Acq On : 18 Jun 2019 21:19
Operator : JC/SP
Sample : K3324-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: Jun 19 06:41:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM061719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 04:35:38 2019
Response via : Initial Calibration

