

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061919\
 Data File : VU032851.D
 Acq On : 19 Jun 2019 12:23
 Operator : JC/SP
 Sample : K3413-25
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GALH1

Quant Time: Jun 20 06:06:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 20 05:14:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	1199	5.00	ug/L	0.01
28) Chlorobenzene-d5	9.09	117	1530	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	1424	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.40	65	28977	371.68	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 7433.60%#			
7) Chloroethane-d5	1.68	69	24782	371.88	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 7437.60%#			
11) 1,1-Dichloroethene-d2	2.27	63	47590	300.90	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 6018.00%#			
20) 2-Butanone-d5	4.19	46	99062	4553.27	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery = 9106.54%#			
24) Chloroform-d	4.64	84	60998	432.51	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 8650.20%#			
26) 1,2-Dichloroethane-d4	5.31	65	35533	487.12	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 9742.40%#			
32) Benzene-d6	5.33	84	119373	313.85	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 6277.00%#			
36) 1,2-Dichloropropane-d6	6.32	67	38158	325.58	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 6511.60%#			
41) Toluene-d8	7.56	98	106809	303.68	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 6073.60%#			
43) trans-1,3-Dichloropropene-	7.85	79	15288	305.15	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 6103.00%#			
46) 2-Hexanone-d5	8.31	63	72523	2888.96	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 5777.92%#			
57) 1,1,2,2-Tetrachloroethane-	10.43	84	32144	325.26	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 6505.20%#			
64) 1,2-Dichlorobenzene-d4	11.85	152	45745	178.59	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 3571.80%#			
Target Compounds						
13) Acetone	2.32	43	17254	1021.809	ug/L	96
42) Toluene	7.63	91	3761	7.481	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061919\
Data File : VU032851.D
Acq On : 19 Jun 2019 12:23
Operator : JC/SP
Sample : K3413-25
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GALH1

Quant Time: Jun 20 06:06:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 20 05:14:15 2019
Response via : Initial Calibration

