

Data File : VU032859.D
Acq On : 19 Jun 2019 15:35
Operator : JC/SP
Sample : K3413-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GALG5

Quant Time: Jun 28 07:21:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 20 05:14:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	96817	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	92347	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	50206	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	27644	4.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.80%
7) Chloroethane-d5	1.68	69	23601	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.80%
11) 1,1-Dichloroethene-d2	2.27	63	45834	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.19	46	95719	54.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.98%
24) Chloroform-d	4.64	84	56994	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.31	65	32957	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.33	84	112044	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.32	67	36551	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.56	98	104255	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
43) trans-1,3-Dichloropropene-	7.85	79	14942	4.94	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.80%
46) 2-Hexanone-d5	8.31	63	73589	48.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.14%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	31410	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.40%
64) 1,2-Dichlorobenzene-d4	11.85	152	45003	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds

					Qvalue
30) Cyclohexane	4.98	56	13914	1.170 ug/L	93
33) Benzene	5.38	78	86626	3.182 ug/L	100
35) Methylcyclohexane	6.39	83	12778	1.014 ug/L #	53
42) Toluene	7.63	91	5872	0.194 ug/L	99
52) Ethylbenzene	9.25	91	10851	0.324 ug/L	98
53) m,p-Xylene	9.37	106	3017	0.238 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU032859.D
Acq On : 19 Jun 2019 15:35
Operator : JC/SP
Sample : K3413-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GALG5

Quant Time: Jun 28 07:21:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 20 05:14:15 2019
Response via : Initial Calibration

