

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061919\
 Data File : VU032877.D
 Acq On : 19 Jun 2019 22:54
 Operator : JC/SP
 Sample : K3413-12
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GALG0

Quant Time: Jun 20 08:42:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 20 07:24:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	91853	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	91869	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48358	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	28076	4.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.00%
7) Chloroethane-d5	1.68	69	24656	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.27	63	45536	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.20%
20) 2-Butanone-d5	4.19	46	102745	61.65	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.30%
24) Chloroform-d	4.64	84	61295	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.40%
26) 1,2-Dichloroethane-d4	5.31	65	36566	6.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	130.80%#
32) Benzene-d6	5.33	84	117755	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
36) 1,2-Dichloropropane-d6	6.32	67	38895	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.56	98	105663	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
43) trans-1,3-Dichloropropene-	7.85	79	15296	5.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.60%
46) 2-Hexanone-d5	8.31	63	77257	51.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.50%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	33138	5.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	46467	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds


					Ovalue
13) Acetone	2.32	43	3376	2.610	ug/L 97
42) Toluene	7.63	91	8063	0.267	ug/L 96
47) Tetrachloroethene	8.22	164	51035	8.324	ug/L 95
53) m,p-Xylene	9.38	106	1789	0.142	ug/L 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061919\
Data File : VU032877.D
Acq On : 19 Jun 2019 22:54
Operator : JC/SP
Sample : K3413-12
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GALG0

Quant Time: Jun 20 08:42:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 20 07:24:34 2019
Response via : Initial Calibration





#13

Acetone

Concen: 2.610 ug/L

RT: 2.32 min Scan# 382

Delta R.T. 0.00 min

Lab File: VU032877.D

Acq: 19 Jun 2019 22:54

Instrument :

MSVOA_U

ClientSampleId :


GALG0

Tot Ion: 43 Resp: 3376

Ion Ratio Lower Upper

43 100

58 31.8 0.0 60.8





Abundance Ion 43.05 (42.75 to 43.75): VU032860.D (-364) (-)

Ion 58.05 (57.75 to 58.75): VU032860.D (-364) (-)

2.32

2.25 2.30 2.35 2.40

Time-->



#42

Toluene

Concen: 0.267 ug/L

RT: 7.63 min Scan# 2035

Delta R.T. 0.00 min

Lab File: VU032877.D


Acq: 19 Jun 2019 22:54

Tgt Ion: 91 Resp: 8063

Ion Ratio Lower Upper

91 100

92 61.2 40.9 75.9




Abundance Ion 91.15 (90.85 to 91.85): VU032860.D (-2020) (-)


Ion 92.15 (91.85 to 92.85): VU032860.D (-2020) (-)

7.63

7.60 7.65

Time-->





#47

Tetrachloroethene

Concen: 8.324 ug/L

RT: 8.22 min Scan# 2217

Delta R.T. -0.00 min

Lab File: VU032877.D

Acq: 19 Jun 2019 22:54

Instrument :

MSVOA_U

ClientSampleId :

GALG0

Tot Ion:164 Resp: 51035


Ion Ratio Lower Upper

164 100

129 92.9 66.1 122.7

131 89.3 65.6 121.8

166 123.8 93.2 173.2






Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V



#53

m,p-Xylene

Concen: 0.142 ug/L

RT: 9.38 min Scan# 2577

Delta R.T. 0.01 min

Lab File: VU032877.D


Acq: 19 Jun 2019 22:54

Tgt Ion:106 Resp: 1789

Ion Ratio Lower Upper

106 100

91 195.4 142.4 264.4



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VU0

