

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061923\
 Data File : VU054590.D
 Acq On : 19 Jun 2023 12:41
 Operator : MD/SY
 Sample : 03258-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCGW8

Quant Time: Jun 20 05:51:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 17 06:10:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	253193	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	242326	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119176	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	52761	3.274	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.400%
7) Chloroethane-d5	1.911	69	74872	4.963	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%
11) 1,1-Dichloroethene-d2	2.563	65	25985	3.611	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	4.624	46	255298	56.352	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.700%
24) Chloroform-d	5.062	84	175031	5.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.698	65	90534	5.514	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
32) Benzene-d6	5.724	84	336900	5.001	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
36) 1,2-Dichloropropane-d6	6.685	67	120129	5.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.894	98	267696	4.478	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	8.177	79	35118	4.489	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.800%
46) 2-Hexanone-d5	8.631	63	179720	54.759	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.520%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	92070	5.488	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	108549	5.400	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061923\
Data File : VU054590.D
Acq On : 19 Jun 2023 12:41
Operator : MD/SY
Sample : 03258-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCGW8

Quant Time: Jun 20 05:51:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 17 06:10:25 2023
Response via : Initial Calibration

