

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061923\
 Data File : VU054591.D
 Acq On : 19 Jun 2023 13:05
 Operator : MD/SY
 Sample : 03237-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YBW76

Quant Time: Jun 20 05:51:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 17 06:10:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	264626	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	258375	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	128354	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	95552	5.674	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 113.400%		
7) Chloroethane-d5	1.911	69	136556	8.660	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 173.200%#		
11) 1,1-Dichloroethene-d2	2.567	65	47978	6.379	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 127.600%#		
20) 2-Butanone-d5	4.621	46	472152	99.715	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 199.440%#		
24) Chloroform-d	5.059	84	324325	9.213	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 184.200%#		
26) 1,2-Dichloroethane-d4	5.698	65	165599	9.650	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 193.000%#		
32) Benzene-d6	5.724	84	617616	8.599	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 172.000%#		
36) 1,2-Dichloropropane-d6	6.685	67	221616	9.239	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 184.800%#		
41) Toluene-d8	7.894	98	503342	7.897	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 158.000%#		
43) trans-1,3-Dichloroprop...	8.177	79	63555	7.619	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 152.400%#		
46) 2-Hexanone-d5	8.627	63	338257	96.662	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 193.320%#		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	168193	9.403	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 188.000%#		
66) 1,2-Dichlorobenzene-d4	12.187	152	199090	9.196	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 184.000%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061923\
Data File : VU054591.D
Acq On : 19 Jun 2023 13:05
Operator : MD/SY
Sample : 03237-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YBW76

Quant Time: Jun 20 05:51:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 17 06:10:25 2023
Response via : Initial Calibration

