

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062019\
 Data File : VU032906.D
 Acq On : 20 Jun 2019 17:38
 Operator : JC/SP
 Sample : K3413-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GALFO

Quant Time: Jun 21 07:48:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 07:36:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	83345	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	83163	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	43327	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	22981	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.68	69	20603	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.27	63	37918	3.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.00%
20) 2-Butanone-d5	4.18	46	98579	65.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.36%#
24) Chloroform-d	4.64	84	54555	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.20%
26) 1,2-Dichloroethane-d4	5.30	65	32793	6.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	129.40%
32) Benzene-d6	5.33	84	99919	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.32	67	34027	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.56	98	90168	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.85	79	13844	5.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.60%
46) 2-Hexanone-d5	8.31	63	70384	51.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	30981	5.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.40%
64) 1,2-Dichlorobenzene-d4	11.85	152	40520	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds					Ovalue
13) Acetone	2.32	43	2960	2.522 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062019\
Data File : VU032906.D
Acq On : 20 Jun 2019 17:38
Operator : JC/SP
Sample : K3413-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GALF0

Quant Time: Jun 21 07:48:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 07:36:10 2019
Response via : Initial Calibration

