

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062019\
 Data File : VU032905.D
 Acq On : 20 Jun 2019 17:11
 Operator : JC/SP
 Sample : VU0620WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK06

Quant Time: Jun 21 07:45:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 07:36:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	89280	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	89984	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	46411	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	21749	3.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.00%
7) Chloroethane-d5	1.68	69	19818	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.80%
11) 1,1-Dichloroethene-d2	2.27	63	35819	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	4.18	46	90321	55.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.50%
24) Chloroform-d	4.64	84	51166	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
26) 1,2-Dichloroethane-d4	5.30	65	31159	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.80%
32) Benzene-d6	5.33	84	96245	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.00%
36) 1,2-Dichloropropane-d6	6.32	67	31446	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.56	98	83239	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	7.85	79	12555	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
46) 2-Hexanone-d5	8.31	63	65456	44.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.66%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	28742	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	37946	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062019\
Data File : VU032905.D
Acq On : 20 Jun 2019 17:11
Operator : JC/SP
Sample : VU0620WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK06

Quant Time: Jun 21 07:45:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 07:36:10 2019
Response via : Initial Calibration

