

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062019\
 Data File : VU032925.D
 Acq On : 21 Jun 2019 01:16
 Operator : JC/SP
 Sample : K3413-27
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 21 08:59:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 07:36:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	82870	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	83967	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	41340	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	20687	3.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.80%
7) Chloroethane-d5	1.68	69	19680	4.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.40%
11) 1,1-Dichloroethene-d2	2.27	63	35148	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	4.19	46	90639	60.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.56%
24) Chloroform-d	4.64	84	51114	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.30	65	30675	6.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.60%
32) Benzene-d6	5.33	84	92120	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.32	67	32261	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.56	98	80584	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	7.85	79	11583	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.31	63	62899	45.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.32%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	29010	5.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	35353	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062019\
Data File : VU032925.D
Acq On : 21 Jun 2019 01:16
Operator : JC/SP
Sample : K3413-27
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 21 08:59:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 07:36:10 2019
Response via : Initial Calibration

