

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
 Data File : VU049304.D
 Acq On : 20 Jun 2022 14:38
 Operator : SY/MD
 Sample : N3395-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0159

Quant Time: Jun 21 01:11:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 21 01:08:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	255386	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	244078	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	120948	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	53002	33.689	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.380%
7) Chloroethane-d5	1.912	69	47295	40.942	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.880%
11) 1,1-Dichloroethene-d2	2.568	63	78782	26.537	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.080%#
21) 2-Butanone-d5	4.623	46	175946	102.001	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.000%
24) Chloroform-d	5.063	84	138950	41.400	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.800%
26) 1,2-Dichloroethane-d4	5.703	65	95642	43.985	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.980%
32) Benzene-d6	5.726	84	257729	42.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.160%
36) 1,2-Dichloropropane-d6	6.690	67	97081	47.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.420%
41) Toluene-d8	7.896	98	222484	41.682	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.360%
43) trans-1,3-Dichloroprop...	8.179	79	41175	43.016	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.040%
47) 2-Hexanone-d5	8.632	63	106835	100.326	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.330%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	155358	45.717	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.440%
66) 1,2-Dichlorobenzene-d4	12.192	152	94826	45.060	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
Data File : VU049304.D
Acq On : 20 Jun 2022 14:38
Operator : SY/MD
Sample : N3395-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0159

Quant Time: Jun 21 01:11:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 21 01:08:31 2022
Response via : Initial Calibration

