

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
 Data File : VU049308.D
 Acq On : 20 Jun 2022 16:12
 Operator : SY/MD
 Sample : N3395-17
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0164

Quant Time: Jun 21 01:12:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 21 01:08:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	243351	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	232141	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	115929	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	67063	44.734	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.460%
7) Chloroethane-d5	1.912	69	55488	50.410	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.820%
11) 1,1-Dichloroethene-d2	2.565	63	95055	33.602	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.200%
21) 2-Butanone-d5	4.620	46	188816	114.875	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.880%
24) Chloroform-d	5.063	84	154946	48.450	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.900%
26) 1,2-Dichloroethane-d4	5.703	65	105671	51.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.000%
32) Benzene-d6	5.729	84	292722	50.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.700%
36) 1,2-Dichloropropane-d6	6.690	67	108921	56.286	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.580%
41) Toluene-d8	7.896	98	263193	51.844	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.680%
43) trans-1,3-Dichloroprop...	8.179	79	47648	52.338	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.680%
47) 2-Hexanone-d5	8.632	63	116949	115.472	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.470%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	165998	51.360	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.720%
66) 1,2-Dichlorobenzene-d4	12.192	152	105339	52.223	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
Data File : VU049308.D
Acq On : 20 Jun 2022 16:12
Operator : SY/MD
Sample : N3395-17
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0164

Quant Time: Jun 21 01:12:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 21 01:08:31 2022
Response via : Initial Calibration

