

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
 Data File : VU049316.D
 Acq On : 20 Jun 2022 19:20
 Operator : SY/MD
 Sample : N3396-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jun 21 01:14:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 21 01:08:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	245300	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	237874	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	117575	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	67513	44.676	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.360%
7) Chloroethane-d5	1.913	69	56993	51.366	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.740%
11) 1,1-Dichloroethene-d2	2.565	63	96572	33.867	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.740%
21) 2-Butanone-d5	4.620	46	189158	114.169	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.170%
24) Chloroform-d	5.064	84	156938	48.683	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.360%
26) 1,2-Dichloroethane-d4	5.700	65	104425	49.999	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.000%
32) Benzene-d6	5.726	84	295593	50.113	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.220%
36) 1,2-Dichloropropane-d6	6.690	67	110745	55.849	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.700%
41) Toluene-d8	7.896	98	266940	51.315	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.640%
43) trans-1,3-Dichloroprop...	8.179	79	46126	49.445	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.900%
47) 2-Hexanone-d5	8.633	63	114651	110.474	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.470%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	172132	51.974	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.940%
66) 1,2-Dichlorobenzene-d4	12.192	152	106579	52.098	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062022\
Data File : VU049316.D
Acq On : 20 Jun 2022 19:20
Operator : SY/MD
Sample : N3396-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jun 21 01:14:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 21 01:08:31 2022
Response via : Initial Calibration

