

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062023\
 Data File : VU054620.D
 Acq On : 20 Jun 2023 19:11
 Operator : MD/SY
 Sample : 03237-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YBW80

Quant Time: Jun 21 04:04:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 21 04:02:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	321852	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	319150	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	146121	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.602	65	129630	6.329	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	126.600%
7) Chloroethane-d5	1.914	69	100386	5.235	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.567	65	51207	5.597	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.000%
20) 2-Butanone-d5	4.631	46	270909	47.041	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.080%
24) Chloroform-d	5.059	84	191186	4.465	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.699	65	98142	4.702	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	5.724	84	391146	4.409	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.686	67	134504	4.540	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.800%
41) Toluene-d8	7.895	98	327856	4.164	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.200%
43) trans-1,3-Dichloroprop...	8.174	79	45843	4.449	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.631	63	178769	41.358	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.720%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	94128	4.260	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	112989	4.584	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062023\
Data File : VU054620.D
Acq On : 20 Jun 2023 19:11
Operator : MD/SY
Sample : 03237-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YBW80

Quant Time: Jun 21 04:04:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 21 04:02:18 2023
Response via : Initial Calibration

