

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024806.D
 Acq On : 21 Jun 2018 13:15
 Operator : MD/SY
 Sample : J3556-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SRI3-TB1

Quant Time: Jun 22 04:01:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	201038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305368	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	289113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	157136	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	142594	43.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.90%	
35) Dibromofluoromethane	4.89	113	120233	47.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.88%	
50) Toluene-d8	7.57	98	447160	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
62) 4-Bromofluorobenzene	10.31	95	165863	45.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.38%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024806.D
Acq On : 21 Jun 2018 13:15
Operator : MD/SY
Sample : J3556-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SRI3-TB1

Quant Time: Jun 22 04:01:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

