

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062118\
 Data File : VU024812.D
 Acq On : 21 Jun 2018 15:40
 Operator : MD/SY
 Sample : J3517-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S39-0182

Quant Time: Jun 22 04:26:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	204197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	309841	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	293952	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	157575	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	144081	43.22	ug/l	0.00
Spiked Amount						
			Recovery	=	86.44%	
35) Dibromofluoromethane	4.89	113	120973	47.04	ug/l	0.00
Spiked Amount						
			Recovery	=	94.08%	
50) Toluene-d8	7.57	98	448544	47.82	ug/l	0.00
Spiked Amount						
			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	10.31	95	167531	44.99	ug/l	0.00
Spiked Amount						
			Recovery	=	89.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	11570	6.46	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062118\
Data File : VU024812.D
Acq On : 21 Jun 2018 15:40
Operator : MD/SY
Sample : J3517-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S39-0182

Quant Time: Jun 22 04:26:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

